

# **Requirement Specification for SARS-Cov-2 Antigen test. (Corona Quick test)**

## 1. Requirements

### 1.1. Description of the requirement

DALO needs to acquire SARS-CoV-2 Antigen test, also known as Corona Quick or Corona Rapid Ag test, to perform testing of soldiers national and abroad.

### 1.2. Description and definitions

The requirement specification, cf. section 1.4, describes all the requirements for the acquisition and consists of six columns with the following information:

"#"	ID number
"Requirement"	Requirement description
"Classification"	The classification of the requirement as further described in section 1.3
"DALO remarks"	Further information regarding the requirement
"Requirement compliance"	The tenderer's indication of compliance (YES or NO)
"Tender description"	Requirements regarding the tenderer's compliance description

### 1.3. Classification

All requirements are mandatory requirements (SHALL) and shall be fulfilled by the tenderer. If just one of the mandatory requirements is not fulfilled, the tenderer's tender will not be taken into further consideration.

#### 1.4. Requirement and response sheet

#	Requirement	Classification	DALO remarks	To be filled out by the tenderer		
				Requirement compliance (tick a box)		Tenderer's description
				YES	NO	
1	The test shall have a relative sensitivity of min. 96,50%	SHALL				
2	The test shall have a relative specificity of min. 99,20%	SHALL				
3	The test shall have an accuracy of minimum 98.30%	SHALL				
4	The test shall produce a result of maximum 30 minutes	SHALL	The time line is from the tests swab is introduced to the test material, and until the result is produced			

<b>5</b>	The test shall be delivered in boksos of maksimum 25 individually packed tests	SHALL				
<b>6</b>	The test shall have a minimum lifetime of 18 month from production date.	SHALL				
<b>7</b>	The tender shall deliver 16000 tests within 10 workdays.	SHALL	DALO kan accept partial delivery in batches of minimum 4000 tests.			